



## **DECLARATION OF CONFORMITY**

**FULL QUALITY ASSURANCE SYSTEM** 

Product Name: Angled Handpiece

Model Numbers: 16-0100

**GMDN Code:** Handpiece, surgical [17949]

**Applied Standards:** • ISO 15883-1:2006 • BS EN ISO 15223-1:2016

■ ISO 10993-1:2009

ANSI AAMI
 BS EN ISO 14971:2012
 ST81:2004/(R)2010
 BS EN 1041:2008+A1:2013

■ ISO 13485:2003

■ ISO 3964:2016

**Manufacturer:** ELLIQUENCE, LLC Phone: (516) 277-9000

2455 Grand Avenue Fax: (516) 277-9001

Baldwin, New York 11510 Email: quality@elliquence.com

USA

**Serial Number:** 

I certify, in my capacity as CEO of Elliquence, LLC that all relevant requirements of the following recognized consensus standards are met with respect to the manufacture of Elliquence products. All records relating to the manufacturers' compliance with these standards will be maintained for a period of two years or for the expected design life of the device, whichever is longer [21 USC 514(c)(3)(C)]. It is understood that these records may be subject to review during an inspection to assess conformance to the Quality System Regulation (21 CFR 820).

November 2, 2020
Alan Ellman, CEO
Date

elliquence, LLC.