

DECLARATION OF CONFORMITY

We, of the company

NOUVAG AG
St.Gallerstrasse 25
CH-9403 Goldach
Switzerland

With the following EU Authorized Representative



Nouvag Dental- und Medizintechnik GmbH
Schulthaißstr. 15, 78462, Konstanz, DE
SRN: DE-AR-000005643

declare on our own responsibility that the medical device

Basic UDI-DI: ++ENOU100021LD

HighSurg 30

<i>REF</i>	<i>Description</i>	<i>MDD classification (Rule)</i>	<i>UMDNS</i>	<i>GMDN</i>
3360 3360tek	Control unit HighSurg 30 100V; 115V; 230V	Ila (9)	16-868	47859

meets all the provisions of the Directive 93/42/EEC which applies to it.

Conformity assessment procedure

93/42/EEC, Appendix II

Validity of the declaration of conformity

26. Mai 2024 (Reg.-Nr. HD 1709948-1)

Notified body

TÜV Rheinland LGA Products GmbH
Tillystrasse 2
DE-90431 Nürnberg



Goldach, 16.11.2021

Place, date


C. Micheletti, PRRC